

**medax**

**MED-S**

**BONE MARROW ASPIRATION SYSTEM**



Bone Marrow aspiration system “XS”, also available in the explant/transplant version “ES”.

- Coloured-coded.
- Luer Lock connection for syringe.
- Triple sharpened tip cannula to allow a non traumatic, fast and easy penetration into the bone marrow cavity.
- Adjustable depth setter.
- Sterilized by ETO, shelf life 5 years.

**medax**  
medical devices

Your partner in Biopsy and Special Needles

# MED-S

## BONE MARROW ASPIRATION SYSTEM

### Ordering information:

GAUGE	COLOUR	LENGTH (MM)							
14	● Green	14 020	14 030	14 035	14 040	14 045	14 050	14 055	14 060
15	● Blue	15 020	15 030	15 035	15 040	15 045	15 050	15 055	15 060
16	○ White	16 020	16 030	16 035	16 040	16 045	16 050	16 055	16 060
18	● Pink	18 020	18 030	18 035	18 040	18 045	18 050	18 055	18 060

GAUGE	COLOUR	LENGTH (MM)								
14	● Green	14 065	14 070	14 075	14 080	14 085	14 090	14 095	14 100	14 150
15	● Blue	15 065	15 070	15 075	15 080	15 085	15 090	15 095	15 100	15 150
16	○ White	16 065	16 070	16 075	16 080	16 085	16 090	16 095	16 100	16 150
18	● Pink	18 065	18 070	18 075	18 080	18 085	18 090	18 095	18 100	18 150

Different sizes on demand

### Ordering information:

MED-S ULTRA BEVEL TIP, WITHOUT WINGS	X	S	GAUGE	LENGTH	0	0
MED-S BEVEL TIP, WITHOUT WINGS	X	S	GAUGE	LENGTH	0	1
MED-S ULTRA BEVEL TIP, WITH WINGS	X	S	GAUGE	LENGTH	1	0
MED-S BEVEL TIP, WITH WINGS	X	S	GAUGE	LENGTH	1	1
MED-S EXPLANT/TRANSPLANT VERSION WITH WINGS	E	S	GAUGE	LENGTH	1	1

Single box: 10 PCS.

**Medax Srl Unipersonale**

**Headquarters:** Via S. Pertini, 4 • 41039 • San Possidonio (MO) • Italy  
 Company direct No. : +39 0535 1812757 • Fax No : +39 0535 1812744  
 email: customercare@medax.it • PEC: medax@legalmail.it • www.medax.it

**Registered Office:** Via R. Piva, 1/A • 46025 • Poggio Rusco (MN) • Italy  
 Vat N. /Fiscal Code N. Iscriz. Reg. Impr.: MN 02669860369 • N. REA: MN 233527, MO 403036 • Capitale Sociale Euro 100.011,00 i.v.



Medax's quality management system is certified to ISO 13485: 2016 standards.  
 In accordance with the requirements of the medical device 93/42/EEC directive and its relevant updates.  
 All products undergo intensive clinical testing and are fully EC and FDA approved.