

medax

MEDONE ULTRA

PROGRAMMABLE AUTOMATIC BIOPSY SYSTEM



Programmable automatic disposable biopsy system.

- One hand usage.
- 25 mm penetration depth.
- 20 mm sample notch.
- Automatic or sequential firing modes.
- Echogenic treatment for ultrasound guided procedures.
- Centimeter markings.
- Can be used with CT introducer coaxial needle.
- Sterilized by ETO, shelf life 5 years






medax
medical devices

Your partner in Biopsy and Special Needles

MEDaX

MEDONE ULTRA

PROGRAMMABLE AUTOMATIC BIOPSY SYSTEM

GAUGE	COLOUR	LENGTH (MM)					
12	 Light blue	12 060	12 100	12 120	12 160	12 200	
14	 Green	14 060	14 100	14 120	14 160	14 200	
16	 White	16 060	16 100	16 120	16 160	16 200	
18	 Pink	18 060	18 100	18 120	18 160	18 200	18 250
20	 Yellow	20 060	20 100	20 120	20 160	20 200	

Different sizes on demand

Ordering information:

MEDONE ULTRA

M	U	GAUGE	LENGTH	0	0
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

INTRODUCER COAXIAL NEEDLE FOR MEDONE ULTRA

C	T	GAUGE	LENGTH	0	0
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

MEDONE ULTRA WITH COAXIAL NEEDLE IN THE SAME PACKAGE

M	U	GAUGE	LENGTH	C	0
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

MEDONE ULTRA WITH COAXIAL NEEDLE INCLUDING BLUNT TIP STYLET IN THE SAME PACKAGE

M	U	GAUGE	LENGTH	K	0
		<input type="text"/>	<input type="text"/>	<input type="text"/>	<input type="text"/>

Single box: 5 PCS.

Medax Srl Unipersonale

Headquarters: Via S. Pertini, 4 • 41039 • San Possidonio (MO) • Italy
 Company direct No. : +39 0535 1812757 • Fax No : +39 0535 1812744
 email: customercare@medax.it • PEC: medax@legalmail.it • www.medax.it

Registered Office: Via R. Piva, 1/A • 46025 • Poggio Rusco (MN) • Italy
 Vat N. /Fiscal Code N. Iscriz. Reg. Impr.: MN 02669860369 • N. REA: MN 233527, MO 403036 • Capitale Sociale Euro 100.011,00 i.v.



Medax's quality management system is certified to ISO 13485:2016 standards.
 In accordance with the requirements of the medical device 93/42/EEC directive and its relevant updates.
 All products undergo intensive clinical testing and are fully EC and FDA approved.