

medax

MEDMAG

AUTOMATIC REUSABLE BIOPSY SYSTEM



medax

UNIMAG

MEDMAG COMPATIBLE BIOPSY NEEDLE
ALSO USABLE WITH ARGON MEDICAL
PRO-MAG™ ULTRA 2.5 BIOPSY SYSTEM

Automatic reusable biopsy system usable with disposable needle UNIMAG.

MEDMAG

- Light Weight.
- 25mm penetration depth.
- Safety switch manually activate.
- Double firing trigger in proximal and distal positions.

UNIMAG

- MEDMAG compatible needle.
- Echogenic treatment for ultrasound guided procedures.
- Centimeter markings.
- Also available with Trocar tip.
- Can be used with CU introducer coaxial needle.
- Sterilized by ETO, shelf life 5 years.






medax
medical devices

Your partner in Biopsy and Special Needles

medax

MEDMAG

AUTOMATIC REUSABLE BIOPSY SYSTEM

GAUGE	COLOUR	LENGTH (MM)							
12	 Light blu	12 080	12 100	12 120	12 160	12 180	12 200	12 250	12 300
14	 Green	14 080	14 100	14 120	14 160	14 180	14 200	14 250	14 300
16	 White	16 080	16 100	16 120	16 160	16 180	16 200	16 250	16 300
18	 Pink	18 080	18 100	18 120	18 160	18 180	18 200	18 250	18 300
20	 Yellow	20 080	20 100	20 120	20 160	20 180	20 200	20 250	20 300

Different sizes on demand

Ordering information:

MEDMAG

M E D M A G - 0 0 / N

UNIMAG NORMAL TIP

U	M	GAUGE		LENGTH			0	0

UNIMAG TROCAR TIP

U	M	GAUGE		LENGTH			0	1

INTRODUCER COAXIAL NEEDLE FOR UNIMAG

C	U	GAUGE		LENGTH			0	0

UNIMAG WITH COAXIAL NEEDLE IN THE SAME PACKAGE

U	M	GAUGE		LENGTH			C	0

Single box: 10 PCS.

Medax Srl Unipersonale

Headquarters: Via S. Pertini, 4 • 41039 • San Possidonio (MO) • Italy
 Company direct No. : +39 0535 1812757 • Fax No : +39 0535 1812744
 email: customercare@medax.it • PEC: medax@legalmail.it • www.medax.it

Registered Office: Via R. Piva, 1/A • 46025 • Poggio Rusco (MN) • Italy
 Vat N. /Fiscal Code N. Iscriz. Reg. Impr.: MN 02669860369 • N. REA: MN 233527, MO 403036 • Capitale Sociale Euro 100.011,00 i.v.



Medax's quality management system is certified to ISO 13485:2016 standards.
 In accordance with the requirements of the medical device 93/42/EEC directive and its relevant updates.
 All products undergo intensive clinical testing and are fully EC and FDA approved.